

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041823\
 Data File : VX035131.D
 Acq On : 18 Apr 2023 18:24
 Operator : JC/MD
 Sample : 02378-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMK0

Quant Time: Apr 19 00:56:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 19 00:53:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	492945	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	428287	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	188712	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	114232	35.939	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.880%
7) Chloroethane-d5	1.666	69	86275	37.564	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.120%
11) 1,1-Dichloroethene-d2	2.307	65	56749	37.314	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.620%
21) 2-Butanone-d5	4.459	46	165520	78.144	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.140%
24) Chloroform-d	5.056	84	229646	38.365	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.740%
26) 1,2-Dichloroethane-d4	5.958	65	152654	39.713	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.420%
32) Benzene-d6	5.977	84	492226	40.009	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.020%
36) 1,2-Dichloropropane-d6	7.306	67	149350	39.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.480%
41) Toluene-d8	8.647	98	438951	39.038	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.080%#
43) trans-1,3-Dichloroprop...	8.952	79	52371	31.887	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.780%
47) 2-Hexanone-d5	9.385	63	118823	78.734	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.730%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	180729	39.268	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.540%
66) 1,2-Dichlorobenzene-d4	12.323	152	142882	39.071	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.140%#
Target Compounds						
					Qvalue	
13) Acetone	2.386	43	5141	3.272	ug/L	95
51) Chlorobenzene	10.080	112	740472	83.524	ug/L	100
64) 1,3-Dichlorobenzene	11.969	146	55994	9.568	ug/L	99
65) 1,4-Dichlorobenzene	12.043	146	285826	47.729	ug/L	98
67) 1,2-Dichlorobenzene	12.335	146	127895	21.553	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041823\
Data File : VX035131.D
Acq On : 18 Apr 2023 18:24
Operator : JC/MD
Sample : 02378-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMK0

Quant Time: Apr 19 00:56:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 19 00:53:18 2023
Response via : Initial Calibration

