

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041824\
 Data File : VX041159.D
 Acq On : 18 Apr 2024 13:21
 Operator : JC/MD
 Sample : P2129-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-041724

Quant Time: Apr 19 03:31:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	233523	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	343855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	122511	48.592	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.180%
35) Dibromofluoromethane	5.385	113	105545	45.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	8.647	98	394902	49.082	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	139246	45.591	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
20) Methylene Chloride	2.788	84	3959	1.751	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041824\
Data File : VX041159.D
Acq On : 18 Apr 2024 13:21
Operator : JC/MD
Sample : P2129-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-041724

Quant Time: Apr 19 03:31:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

