

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041824\
 Data File : VX041160.D
 Acq On : 18 Apr 2024 13:44
 Operator : JC/MD
 Sample : PB160339TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160339TB

Quant Time: Apr 19 03:32:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	231928	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	340776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	120249	48.023	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.040%
35) Dibromofluoromethane	5.385	113	105084	46.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	392299	49.199	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.400%
62) 4-Bromofluorobenzene	11.079	95	136558	45.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	37235	35.378	ug/l	# 89
18) Methyl Acetate	2.709	43	3031	1.154	ug/l	94
20) Methylene Chloride	2.788	84	6854	3.052	ug/l	85
43) Isopropyl Acetate	6.348	43	128433	25.129	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041824\
Data File : VX041160.D
Acq On : 18 Apr 2024 13:44
Operator : JC/MD
Sample : PB160339TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160339TB

Quant Time: Apr 19 03:32:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

