

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
 Data File : VX045846.D
 Acq On : 18 Apr 2025 10:41
 Operator : JC/MD
 Sample : PB167623TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167623TB

Quant Time: Apr 19 00:56:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132403	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63499	52.417	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.840%
35) Dibromofluoromethane	5.385	113	46073	49.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.100%
50) Toluene-d8	8.647	98	164194	50.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	11.079	95	62218	52.094	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.180%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	24169	48.587	ug/l	99
18) Methyl Acetate	2.709	43	1468	1.282	ug/l #	84
20) Methylene Chloride	2.782	84	2226	2.390	ug/l	92
43) Isopropyl Acetate	6.342	43	11096	4.580	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
Data File : VX045846.D
Acq On : 18 Apr 2025 10:41
Operator : JC/MD
Sample : PB167623TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167623TB

Quant Time: Apr 19 00:56:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

