

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
 Data File : VX045853.D
 Acq On : 18 Apr 2025 13:23
 Operator : JC/MD
 Sample : Q1826-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-FLOOR

Quant Time: Apr 18 23:38:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69352	53.028	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.060%	
35) Dibromofluoromethane	5.385	113	49012	48.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.780%	
50) Toluene-d8	8.647	98	176111	50.329	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	11.079	95	68034	53.378	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.760%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	90464	168.450	ug/l	100
20) Methylene Chloride	2.788	84	3035	3.019	ug/l	96
25) 2-Butanone	4.586	43	4207	5.352	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
Data File : VX045853.D
Acq On : 18 Apr 2025 13:23
Operator : JC/MD
Sample : Q1826-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-FLOOR

Quant Time: Apr 18 23:38:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

