

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
 Data File : VX045854.D
 Acq On : 18 Apr 2025 13:47
 Operator : JC/MD
 Sample : Q1828-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT4643

Quant Time: Apr 18 23:39:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	70139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	131851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70338	54.838	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.680%	
35) Dibromofluoromethane	5.379	113	50744	50.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.460%	
50) Toluene-d8	8.647	98	177454	50.831	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.660%	
62) 4-Bromofluorobenzene	11.079	95	67427	53.025	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.040%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30684	58.258	ug/l	100
18) Methyl Acetate	2.703	43	1853	1.528	ug/l	95
20) Methylene Chloride	2.788	84	2734	2.773	ug/l	87
43) Isopropyl Acetate	6.342	43	7327	2.841	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
Data File : VX045854.D
Acq On : 18 Apr 2025 13:47
Operator : JC/MD
Sample : Q1828-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT4643

Quant Time: Apr 18 23:39:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

