

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
 Data File : VX045855.D
 Acq On : 18 Apr 2025 14:10
 Operator : JC/MD
 Sample : Q1828-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT5417

Quant Time: Apr 18 23:39:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143735	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	134250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70614	53.101	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.200%
35) Dibromofluoromethane	5.379	113	51195	50.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
50) Toluene-d8	8.647	98	179686	50.481	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	68908	53.147	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.300%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	26786	49.054	ug/l	99
18) Methyl Acetate	2.703	43	1831	1.456	ug/l	95
20) Methylene Chloride	2.782	84	2852	2.790	ug/l #	83
43) Isopropyl Acetate	6.342	43	9952	3.784	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
Data File : VX045855.D
Acq On : 18 Apr 2025 14:10
Operator : JC/MD
Sample : Q1828-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT5417

Quant Time: Apr 18 23:39:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

