

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
 Data File : VX045860.D
 Acq On : 18 Apr 2025 16:05
 Operator : JC/MD
 Sample : Q1832-30
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-714-COMP-05

Quant Time: Apr 18 23:42:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	74763	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	146494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72655	53.141	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.280%
35) Dibromofluoromethane	5.379	113	51696	49.738	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.480%
50) Toluene-d8	8.647	98	184851	50.954	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.900%
62) 4-Bromofluorobenzene	11.079	95	71142	53.837	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.680%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	26745	47.638	ug/l	98
18) Methyl Acetate	2.703	43	1971	1.525	ug/l	94
20) Methylene Chloride	2.788	84	3237	3.080	ug/l #	94
43) Isopropyl Acetate	6.336	43	10701	3.992	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041825\
Data File : VX045860.D
Acq On : 18 Apr 2025 16:05
Operator : JC/MD
Sample : Q1832-30
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-714-COMP-05

Quant Time: Apr 18 23:42:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

