

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041919\
 Data File : VX009067.D
 Acq On : 19 Apr 2019 13:57
 Operator : JC/SP
 Sample : K2449-12DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW309-190416DL

Quant Time: Apr 20 01:20:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	239077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387397	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	334955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	141095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160647	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
35) Dibromofluoromethane	5.50	113	117220	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
50) Toluene-d8	8.71	98	482980	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
62) 4-Bromofluorobenzene	11.13	95	158238	45.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.42%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	54847	17.665	ug/l	91
44) Trichloroethene	7.21	130	15800	5.456	ug/l	97
64) Tetrachloroethene	9.34	164	90173	35.437	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041919\
Data File : VX009067.D
Acq On : 19 Apr 2019 13:57
Operator : JC/SP
Sample : K2449-12DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW309-190416DL

Quant Time: Apr 20 01:20:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

