

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041919\
 Data File : VX009071.D
 Acq On : 19 Apr 2019 15:30
 Operator : JC/SP
 Sample : K2487-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 20 01:32:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239074	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	388345	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336293	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140867	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160647	47.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.36%	
35) Dibromofluoromethane	5.50	113	117688	47.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.44%	
50) Toluene-d8	8.71	98	480993	48.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.78%	
62) 4-Bromofluorobenzene	11.13	95	157677	44.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041919\
Data File : VX009071.D
Acq On : 19 Apr 2019 15:30
Operator : JC/SP
Sample : K2487-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 20 01:32:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

