

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041919\
 Data File : VX009080.D
 Acq On : 19 Apr 2019 19:01
 Operator : JC/SP
 Sample : K2495-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040RW056-190417

Quant Time: Apr 22 11:06:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	361823	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	305563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128129	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	149272	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
35) Dibromofluoromethane	5.50	113	110149	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	8.71	98	446678	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
62) 4-Bromofluorobenzene	11.13	95	145217	44.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.84%	
Target Compounds						
4) Vinyl Chloride	1.40	62	9843	3.244	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	1677	0.665	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	239877	83.794	ug/l	92
44) Trichloroethene	7.21	130	243118	89.883	ug/l	100
64) Tetrachloroethene	9.34	164	59447	25.609	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041919\
Data File : VX009080.D
Acq On : 19 Apr 2019 19:01
Operator : JC/SP
Sample : K2495-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040RW056-190417

Quant Time: Apr 22 11:06:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

