

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX041919\
 Data File : VX009082.D
 Acq On : 19 Apr 2019 19:48
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 02:26:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.576	0.583	-1.2	82	0.00
3 P	Chloromethane	0.707	0.655	7.4	81	0.00
4 C	Vinyl Chloride	0.688	0.680	1.2#	84	0.00
5 T	Bromomethane	0.447	0.363	18.8	79	0.00
6 T	Chloroethane	0.405	0.415	-2.5	86	-0.01
7 T	Trichlorofluoromethane	0.803	0.813	-1.2	86	0.00
8 T	Diethyl Ether	0.367	0.378	-3.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.521	-1.4	89	0.00
10 T	Methyl Iodide	0.511	0.416	18.6	63	0.00
11 T	Tert butyl alcohol	0.156	0.171	-9.6	93	0.02
12 CM	1,1-Dichloroethene	0.536	0.517	3.5#	84	0.00
13 T	Acrolein	0.135	0.129	4.4	78	0.00
14 T	Allyl chloride	1.170	1.262	-7.9	90	0.00
15 T	Acrylonitrile	0.397	0.441	-11.1	95	0.00
16 T	Acetone	0.365	0.389	-6.6	89	0.00
17 T	Carbon Disulfide	1.664	1.428	14.2	78	0.00
18 T	Methyl Acetate	0.912	1.052	-15.4	104	0.00
19 T	Methyl tert-butyl Ether	1.911	2.058	-7.7	92	0.00
20 T	Methylene Chloride	0.625	0.626	-0.2	91	0.00
21 T	trans-1,2-Dichloroethene	0.572	0.567	0.9	87	0.00
22 T	Diisopropyl ether	2.273	2.492	-9.6	94	0.00
23 T	Vinyl Acetate	1.998	2.240	-12.1	92	0.00
24 P	1,1-Dichloroethane	1.141	1.214	-6.4	93	0.00
25 T	2-Butanone	0.584	0.657	-12.5	93	0.00
26 T	2,2-Dichloropropane	0.855	0.809	5.4	81	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.668	-2.9	90	0.00
28 T	Bromochloromethane	0.529	0.532	-0.6	94	0.00
29 T	Tetrahydrofuran	0.391	0.438	-12.0	96	0.00
30 C	Chloroform	1.021	1.098	-7.5#	93	0.00
31 T	Cyclohexane	1.122	1.138	-1.4	87	0.00
32 T	1,1,1-Trichloroethane	0.846	0.889	-5.1	89	0.00
33 S	1,2-Dichloroethane-d4	0.705	0.751	-6.5	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.321	0.332	-3.4	91	0.00
36 T	1,1-Dichloropropene	0.517	0.494	4.4	87	0.00
37 T	Ethyl Acetate	0.665	0.739	-11.1	94	0.00
38 T	Carbon Tetrachloride	0.438	0.443	-1.1	89	0.00
39 T	Methylcyclohexane	0.630	0.600	4.8	84	0.00
40 TM	Benzene	1.526	1.524	0.1	89	0.00
41 T	Methacrylonitrile	0.382	0.409	-7.1	93	0.00
42 TM	1,2-Dichloroethane	0.540	0.556	-3.0	92	0.00
43 T	Isopropyl Acetate	1.072	1.152	-7.5	93	0.00
44 TM	Trichloroethene	0.374	0.353	5.6	86	0.00
45 C	1,2-Dichloropropane	0.423	0.445	-5.2#	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX041919\
 Data File : VX009082.D
 Acq On : 19 Apr 2019 19:48
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 02:26:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.246	0.249	-1.2	90	0.00
47 T	Bromodichloromethane	0.483	0.505	-4.6	91	0.00
48 T	Methyl methacrylate	0.530	0.568	-7.2	92	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	88	0.00
50 S	Toluene-d8	1.280	1.283	-0.2	87	0.00
51 T	4-Methyl-2-Pentanone	0.683	0.739	-8.2	94	0.00
52 CM	Toluene	0.914	0.888	2.8#	87	0.00
53 T	t-1,3-Dichloropropene	0.556	0.572	-2.9	86	0.00
54 T	cis-1,3-Dichloropropene	0.623	0.636	-2.1	88	0.00
55 T	1,1,2-Trichloroethane	0.362	0.374	-3.3	91	0.00
56 T	Ethyl methacrylate	0.650	0.676	-4.0	89	0.00
57 T	1,3-Dichloropropane	0.645	0.667	-3.4	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.350	0.361	-3.1	89	0.00
59 T	2-Hexanone	0.527	0.559	-6.1	92	0.00
60 T	Dibromochloromethane	0.347	0.364	-4.9	88	0.00
61 T	1,2-Dibromoethane	0.374	0.377	-0.8	88	0.00
62 S	4-Bromofluorobenzene	0.452	0.453	-0.2	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.380	0.362	4.7	83	0.00
65 PM	Chlorobenzene	1.055	1.050	0.5	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.393	-5.6	90	0.00
67 C	Ethyl Benzene	1.929	1.938	-0.5#	86	0.00
68 T	m/p-Xylenes	0.701	0.704	-0.4	86	0.00
69 T	o-Xylene	0.682	0.687	-0.7	87	0.00
70 T	Styrene	1.182	1.197	-1.3	86	0.00
71 P	Bromoform	0.296	0.309	-4.4	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.790	3.943	-4.0	87	0.00
74 T	N-amyl acetate	2.176	2.369	-8.9	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.385	1.471	-6.2	91	0.00
76 T	1,2,3-Trichloropropane	1.221	1.265	-3.6	89	0.00
77 T	Bromobenzene	0.956	0.936	2.1	85	0.00
78 T	n-propylbenzene	4.350	4.504	-3.5	85	0.00
79 T	2-Chlorotoluene	2.621	2.658	-1.4	87	0.00
80 T	1,3,5-Trimethylbenzene	3.180	3.261	-2.5	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.476	0.489	-2.7	80	0.00
82 T	4-Chlorotoluene	3.044	3.077	-1.1	86	0.00
83 T	tert-Butylbenzene	3.053	3.208	-5.1	90	0.00
84 T	1,2,4-Trimethylbenzene	3.237	3.285	-1.5	86	0.00
85 T	sec-Butylbenzene	3.642	3.780	-3.8	87	0.00
86 T	p-Isopropyltoluene	3.348	3.393	-1.3	85	0.00
87 T	1,3-Dichlorobenzene	1.674	1.667	0.4	86	0.00
88 T	1,4-Dichlorobenzene	1.697	1.663	2.0	86	0.00
89 T	n-Butylbenzene	3.118	3.175	-1.8	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX041919\
Data File : VX009082.D
Acq On : 19 Apr 2019 19:48
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 20 02:26:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.552	0.579	-4.9	85	0.00
91 T	1,2-Dichlorobenzene	1.653	1.664	-0.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.315	0.336	-6.7	92	0.00
93 T	1,2,4-Trichlorobenzene	1.193	1.201	-0.7	86	0.00
94 T	Hexachlorobutadiene	0.595	0.594	0.2	85	0.00
95 T	Naphthalene	3.822	4.035	-5.6	90	0.00
96 T	1,2,3-Trichlorobenzene	1.177	1.207	-2.5	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6