

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028161.D
 Acq On : 19 Apr 2022 15:42
 Operator : JC/MD
 Sample : N2390-01
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-1-041422

Quant Time: Apr 19 16:36:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	394212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	655473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	595218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	275061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	236153	45.944	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.880%
35) Dibromofluoromethane	5.391	113	209438	45.562	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.120%
50) Toluene-d8	8.647	98	769796	43.601	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.200%
62) 4-Bromofluorobenzene	11.085	95	270646	40.092	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	80.180%
Target Compounds						Qvalue
10) Methyl Iodide	2.447	142	2337	5.714	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028161.D
Acq On : 19 Apr 2022 15:42
Operator : JC/MD
Sample : N2390-01
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-1-041422

Quant Time: Apr 19 16:36:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

