

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028162.D
 Acq On : 19 Apr 2022 16:05
 Operator : JC/MD
 Sample : N2452-07ME
 Misc : 5.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-15-1-4-GRME

Quant Time: Apr 20 08:15:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	406852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	681903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	634263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	394163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	244059	46.007	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.020%		
35) Dibromofluoromethane	5.391	113	214023	44.755	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	89.500%		
50) Toluene-d8	8.647	98	806257	43.897	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	87.800%		
62) 4-Bromofluorobenzene	11.085	95	341398	48.612	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.220%		
Target Compounds					Qvalue	
31) Cyclohexane	5.452	56	29912	4.628	ug/l	92
39) Methylcyclohexane	7.373	83	193312	24.561	ug/l	97
67) Ethyl Benzene	10.195	91	30305	1.336	ug/l	98
68) m/p-Xylenes	10.305	106	99552	10.787	ug/l	97
69) o-Xylene	10.646	106	42956	4.668	ug/l	98
73) Isopropylbenzene	10.963	105	177322	6.545	ug/l	100
78) n-propylbenzene	11.305	91	251274	8.364	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	1318172	57.318	ug/l	96
83) tert-Butylbenzene	11.719	119	74141	3.157	ug/l	85
84) 1,2,4-Trimethylbenzene	11.756	105	2615931	114.430	ug/l	94
85) sec-Butylbenzene	11.896	105	340019	12.002	ug/l	97
86) p-Isopropyltoluene	12.012	119	102101	4.177	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028162.D
Acq On : 19 Apr 2022 16:05
Operator : JC/MD
Sample : N2452-07ME
Misc : 5.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-15-1-4-GRME

Quant Time: Apr 20 08:15:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

