

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028168.D
 Acq On : 19 Apr 2022 18:33
 Operator : JC/MD
 Sample : VX0419WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0419WBL02

Quant Time: Apr 20 10:35:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	486112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	798804	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	749002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	358323	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	323226	50.995	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.000%
35) Dibromofluoromethane	5.385	113	312134	55.719	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	111.440%
50) Toluene-d8	8.647	98	1116895	51.910	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.820%
62) 4-Bromofluorobenzene	11.079	95	409896	49.824	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028168.D
Acq On : 19 Apr 2022 18:33
Operator : JC/MD
Sample : VX0419WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0419WBL02

Quant Time: Apr 20 10:35:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

