

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028171.D
 Acq On : 19 Apr 2022 20:06
 Operator : JC/MD
 Sample : PB144164ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#01

Quant Time: Apr 20 08:18:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	444938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	745237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	731788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	340712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	260420	44.889	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.780%
35) Dibromofluoromethane	5.385	113	242982	46.492	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.980%
50) Toluene-d8	8.653	98	888388	44.258	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.520%
62) 4-Bromofluorobenzene	11.079	95	333162	43.408	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.820%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	27568	14.241	ug/l	98
20) Methylene Chloride	2.788	84	213573	49.615	ug/l	99
43) Isopropyl Acetate	6.342	43	32041	2.846	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028171.D
Acq On : 19 Apr 2022 20:06
Operator : JC/MD
Sample : PB144164ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#01

Quant Time: Apr 20 08:18:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

