

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028173.D
 Acq On : 19 Apr 2022 20:53
 Operator : JC/MD
 Sample : PB144164ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#03

Quant Time: Apr 20 08:19:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	426511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	729593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	730257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	343074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	258129	46.416	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.840%
35) Dibromofluoromethane	5.385	113	238311	46.576	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.160%
50) Toluene-d8	8.653	98	872943	44.421	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.840%
62) 4-Bromofluorobenzene	11.079	95	333528	44.387	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.780%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	20923	10.983	ug/l	91
20) Methylene Chloride	2.788	84	205419	49.783	ug/l	99
43) Isopropyl Acetate	6.342	43	32304	2.931	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028173.D
Acq On : 19 Apr 2022 20:53
Operator : JC/MD
Sample : PB144164ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#03

Quant Time: Apr 20 08:19:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

