

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028176.D
 Acq On : 19 Apr 2022 22:02
 Operator : JC/MD
 Sample : PB144164ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#06

Quant Time: Apr 20 08:19:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	399770	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	689670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	694921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	318837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	247186	47.421	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.840%		
35) Dibromofluoromethane	5.391	113	222753	46.056	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.120%		
50) Toluene-d8	8.653	98	830586	44.712	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	89.420%		
62) 4-Bromofluorobenzene	11.079	95	312195	43.953	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.900%		
Target Compounds					Qvalue	
16) Acetone	2.379	43	19832	11.123	ug/l	92
20) Methylene Chloride	2.788	84	204431	52.857	ug/l	98
43) Isopropyl Acetate	6.342	43	32232	3.094	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028176.D
Acq On : 19 Apr 2022 22:02
Operator : JC/MD
Sample : PB144164ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#06

Quant Time: Apr 20 08:19:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

