

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028177.D
 Acq On : 19 Apr 2022 22:25
 Operator : JC/MD
 Sample : PB144164ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#07

Quant Time: Apr 20 08:20:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	408532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	703697	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	711241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	325081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	251071	47.134	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.260%
35) Dibromofluoromethane	5.391	113	229268	46.458	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.920%
50) Toluene-d8	8.653	98	846713	44.671	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.340%
62) 4-Bromofluorobenzene	11.079	95	318972	44.012	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.020%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20548	11.297	ug/l	94
20) Methylene Chloride	2.788	84	199500	50.476	ug/l	98
43) Isopropyl Acetate	6.342	43	31105	2.926	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028177.D
Acq On : 19 Apr 2022 22:25
Operator : JC/MD
Sample : PB144164ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#07

Quant Time: Apr 20 08:20:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

