

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028178.D
 Acq On : 19 Apr 2022 22:49
 Operator : JC/MD
 Sample : PB144164ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#08

Quant Time: Apr 20 08:20:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	401930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	689274	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	694882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	317568	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	248207	47.361	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.720%
35) Dibromofluoromethane	5.385	113	223597	46.257	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.520%
50) Toluene-d8	8.653	98	832935	44.864	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.720%
62) 4-Bromofluorobenzene	11.079	95	311385	43.865	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.720%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	18654	10.315	ug/l	95
20) Methylene Chloride	2.788	84	191109	49.147	ug/l	99
43) Isopropyl Acetate	6.342	43	30767	2.955	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028178.D
Acq On : 19 Apr 2022 22:49
Operator : JC/MD
Sample : PB144164ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#08

Quant Time: Apr 20 08:20:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

