

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028179.D
 Acq On : 19 Apr 2022 23:12
 Operator : JC/MD
 Sample : PB144164ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#09

Quant Time: Apr 20 08:20:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	409812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	700401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	709714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	321430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	250534	46.886	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.780%
35) Dibromofluoromethane	5.391	113	230408	46.909	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.820%
50) Toluene-d8	8.653	98	851821	45.152	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	90.300%
62) 4-Bromofluorobenzene	11.079	95	320167	44.385	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.780%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	26955	15.205	ug/l	91
20) Methylene Chloride	2.788	84	196107	49.463	ug/l	97
43) Isopropyl Acetate	6.342	43	30517	2.884	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028179.D
Acq On : 19 Apr 2022 23:12
Operator : JC/MD
Sample : PB144164ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#09

Quant Time: Apr 20 08:20:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

