

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028181.D
 Acq On : 19 Apr 2022 23:58
 Operator : JC/MD
 Sample : PB144164ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144164ZHE#11

Quant Time: Apr 20 08:21:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	397178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	686689	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	697396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	320679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	247553	47.802	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.600%		
35) Dibromofluoromethane	5.391	113	223620	46.436	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.880%		
50) Toluene-d8	8.653	98	834833	45.136	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	90.280%		
62) 4-Bromofluorobenzene	11.079	95	312500	44.187	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.380%		
Target Compounds					Qvalue	
16) Acetone	2.373	43	24760	14.338	ug/l	91
20) Methylene Chloride	2.794	84	201300	52.388	ug/l	97
43) Isopropyl Acetate	6.342	43	30959	2.984	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028181.D
Acq On : 19 Apr 2022 23:58
Operator : JC/MD
Sample : PB144164ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144164ZHE#11

Quant Time: Apr 20 08:21:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

