

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028157.D
 Acq On : 19 Apr 2022 13:56
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX041922

Quant Time: Apr 20 08:14:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.323	0.275	14.9	109	0.00
3 P	Chloromethane	0.375	0.307	18.1	108	0.00
4 C	Vinyl Chloride	0.504	0.447	11.3#	113	0.00
5 T	Bromomethane	0.522	0.465	10.9	108	0.00
6 T	Chloroethane	0.461	0.341	26.0#	109	0.00
7 T	Trichlorofluoromethane	1.112	0.960	13.7	109	0.00
8 T	Diethyl Ether	0.382	0.316	17.3	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.358	11.6	125	0.00
10 T	Methyl Iodide	0.509	0.430	15.5	99	0.00
11 T	Tert butyl alcohol	0.135	0.114	15.6	112	0.00
12 CM	1,1-Dichloroethene	0.384	0.320	16.7#	117	0.00
13 T	Acrolein	0.080	0.078	2.5	102	0.00
14 T	Allyl chloride	0.645	0.587	9.0	113	0.00
15 T	Acrylonitrile	0.266	0.237	10.9	105	0.00
16 T	Acetone	0.215	0.205	4.7	127	0.00
17 T	Carbon Disulfide	1.035	0.906	12.5	112	0.00
18 T	Methyl Acetate	0.580	0.509	12.2	112	0.00
19 T	Methyl tert-butyl Ether	1.602	1.538	4.0	110	0.00
20 T	Methylene Chloride	0.484	0.431	11.0	116	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.491	6.1	112	0.00
22 T	Diisopropyl ether	1.469	1.317	10.3	106	0.00
23 T	Vinyl Acetate	1.283	1.183	7.8	106	0.00
24 P	1,1-Dichloroethane	0.867	0.801	7.6	116	0.00
25 T	2-Butanone	0.386	0.346	10.4	106	0.00
26 T	2,2-Dichloropropane	0.847	0.799	5.7	112	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.583	6.4	112	0.00
28 T	Bromochloromethane	0.372	0.336	9.7	104	0.00
29 T	Tetrahydrofuran	0.252	0.213	15.5	102	0.00
30 C	Chloroform	1.041	0.950	8.7#	108	0.00
31 T	Cyclohexane	0.794	0.747	5.9	111	0.00
32 T	1,1,1-Trichloroethane	0.943	0.890	5.6	110	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.596	8.6	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
35 S	Dibromofluoromethane	0.351	0.338	3.7	111	0.00
36 T	1,1-Dichloropropene	0.457	0.437	4.4	111	0.00
37 T	Ethyl Acetate	0.464	0.417	10.1	104	0.00
38 T	Carbon Tetrachloride	0.516	0.512	0.8	111	0.00
39 T	Methylcyclohexane	0.577	0.570	1.2	114	0.00
40 TM	Benzene	1.339	1.256	6.2	106	0.00
41 T	Methacrylonitrile	0.239	0.219	8.4	106	0.00
42 TM	1,2-Dichloroethane	0.497	0.456	8.2	105	0.00
43 T	Isopropyl Acetate	0.755	0.678	10.2	105	0.00
44 TM	Trichloroethene	0.388	0.373	3.9	111	0.00
45 C	1,2-Dichloropropane	0.328	0.299	8.8#	106	0.00
46 T	Dibromomethane	0.257	0.239	7.0	108	0.00
47 T	Bromodichloromethane	0.494	0.466	5.7	106	0.00
48 T	Methyl methacrylate	0.347	0.316	8.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
 Data File : VX028157.D
 Acq On : 19 Apr 2022 13:56
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX041922

Quant Time: Apr 20 08:14:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	106	0.00
50 S	Toluene-d8	1.347	1.309	2.8	110	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.432	10.9	103	0.00
52 CM	Toluene	0.930	0.889	4.4#	108	0.00
53 T	t-1,3-Dichloropropene	0.525	0.524	0.2	108	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.542	3.9	108	0.00
55 T	1,1,2-Trichloroethane	0.378	0.357	5.6	108	0.00
56 T	Ethyl methacrylate	0.548	0.518	5.5	105	0.00
57 T	1,3-Dichloropropane	0.610	0.570	6.6	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.252	1.9	105	0.00
59 T	2-Hexanone	0.375	0.348	7.2	106	0.00
60 T	Dibromochloromethane	0.412	0.408	1.0	108	0.00
61 T	1,2-Dibromoethane	0.415	0.397	4.3	106	0.00
62 S	4-Bromofluorobenzene	0.515	0.505	1.9	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.391	0.373	4.6	112	0.00
65 PM	Chlorobenzene	1.066	1.007	5.5	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.388	2.3	110	0.00
67 C	Ethyl Benzene	1.788	1.743	2.5#	107	0.00
68 T	m/p-Xylenes	0.728	0.724	0.5	108	0.00
69 T	o-Xylene	0.725	0.708	2.3	107	0.00
70 T	Styrene	1.175	1.195	-1.7	109	0.00
71 P	Bromoform	0.313	0.322	-2.9	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.437	3.145	8.5	108	0.00
74 T	N-amyl acetate	1.188	1.073	9.7	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	0.986	11.6	106	0.00
76 T	1,2,3-Trichloropropane	0.984	0.906	7.9	107	0.00
77 T	Bromobenzene	0.895	0.800	10.6	108	0.00
78 T	n-propylbenzene	3.811	3.624	4.9	109	0.00
79 T	2-Chlorotoluene	2.369	2.170	8.4	108	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.779	4.7	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.310	6.3	106	0.00
82 T	4-Chlorotoluene	2.729	2.549	6.6	108	0.00
83 T	tert-Butylbenzene	2.979	2.813	5.6	109	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.777	4.2	108	0.00
85 T	sec-Butylbenzene	3.594	3.473	3.4	110	0.00
86 T	p-Isopropyltoluene	3.100	3.035	2.1	109	0.00
87 T	1,3-Dichlorobenzene	1.682	1.616	3.9	110	0.00
88 T	1,4-Dichlorobenzene	1.766	1.662	5.9	111	0.00
89 T	n-Butylbenzene	2.546	2.525	0.8	111	0.00
90 T	Hexachloroethane	0.498	0.509	-2.2	110	0.00
91 T	1,2-Dichlorobenzene	1.671	1.572	5.9	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.220	11.3	103	0.00
93 T	1,2,4-Trichlorobenzene	0.994	0.952	4.2	113	0.00
94 T	Hexachlorobutadiene	0.443	0.406	8.4	111	0.00
95 T	Naphthalene	3.310	3.198	3.4	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041922\
Data File : VX028157.D
Acq On : 19 Apr 2022 13:56
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX041922

Quant Time: Apr 20 08:14:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	0.938	3.8	111	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6