

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041923\
 Data File : VX035164.D
 Acq On : 19 Apr 2023 13:10
 Operator : JC/MD
 Sample : 02397-09DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMM1DL

Quant Time: Apr 20 02:55:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 20 02:36:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	481594	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	413438	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	192684	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	138797	44.697	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.400%	
7) Chloroethane-d5	1.672	69	110250	49.134	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	98.260%	
11) 1,1-Dichloroethene-d2	2.307	65	69131	46.527	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.060%	
21) 2-Butanone-d5	4.465	46	209409	101.194	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.190%	
24) Chloroform-d	5.056	84	282691	48.340	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.680%	
26) 1,2-Dichloroethane-d4	5.958	65	188969	50.320	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.640%	
32) Benzene-d6	5.970	84	607284	51.133	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.260%	
36) 1,2-Dichloropropane-d6	7.306	67	189247	52.165	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.340%	
41) Toluene-d8	8.647	98	546358	50.336	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.680%	
43) trans-1,3-Dichloroprop...	8.952	79	71815	45.297	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.600%	
47) 2-Hexanone-d5	9.384	63	146484	100.549	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	100.550%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	224209	50.464	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.920%	
66) 1,2-Dichlorobenzene-d4	12.317	152	191948	51.406	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.820%	
Target Compounds						
					Qvalue	
51) Chlorobenzene	10.079	112	188089	21.978	ug/L	100
64) 1,3-Dichlorobenzene	11.969	146	13519	2.262	ug/L	99
65) 1,4-Dichlorobenzene	12.043	146	61063	9.987	ug/L	97
67) 1,2-Dichlorobenzene	12.335	146	316054	52.164	ug/L	97
70) 1,2,4-trichlorobenzene	13.591	180	66555	19.241	ug/L	99
72) 1,2,3-Trichlorobenzene	13.963	180	14189	4.155	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041923\
Data File : VX035164.D
Acq On : 19 Apr 2023 13:10
Operator : JC/MD
Sample : 02397-09DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMM1DL

Quant Time: Apr 20 02:55:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 20 02:36:32 2023
Response via : Initial Calibration

