

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041923\
 Data File : VX035187.D
 Acq On : 19 Apr 2023 22:25
 Operator : JC/MD
 Sample : 02333-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMG7

Quant Time: Apr 20 05:08:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 20 04:51:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	502454	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	425893	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	190584	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	125854	38.846	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.700%
7) Chloroethane-d5	1.672	69	100508	42.933	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.860%
11) 1,1-Dichloroethene-d2	2.307	65	64093	41.345	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.700%
21) 2-Butanone-d5	4.465	46	224046	103.773	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.770%
24) Chloroform-d	5.056	84	278131	45.586	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.180%
26) 1,2-Dichloroethane-d4	5.958	65	188689	48.159	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.320%
32) Benzene-d6	5.977	84	591369	48.337	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.680%
36) 1,2-Dichloropropane-d6	7.306	67	187412	50.149	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.300%
41) Toluene-d8	8.647	98	517348	46.269	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.540%
43) trans-1,3-Dichloroprop...	8.952	79	67471	41.312	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.620%
47) 2-Hexanone-d5	9.385	63	157097	104.680	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.680%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	234850	51.313	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.620%
66) 1,2-Dichlorobenzene-d4	12.323	152	186304	50.444	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041923\
Data File : VX035187.D
Acq On : 19 Apr 2023 22:25
Operator : JC/MD
Sample : 02333-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMG7

Quant Time: Apr 20 05:08:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 20 04:51:49 2023
Response via : Initial Calibration

