



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041923\
Data File : VX035156.D
Acq On : 19 Apr 2023 08:59
Operator : JC/MD
Sample : BFB624
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB624

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM041823WMA.M
Title : VOC Analysis
Last Update : Thu Apr 20 23:27:37 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.1	16702	PASS
75	95	30	80	50.1	46149	PASS
95	95	100	100	100.0	92123	PASS
96	95	5	9	7.1	6523	PASS
173	174	0.00	2	1.1	746	PASS
174	95	50	120	74.9	68995	PASS
175	174	5	9	7.5	5178	PASS
176	174	95	101	97.0	66896	PASS
177	176	5	9	6.2	4168	PASS

m/z	Abundance
39.90	166.0
44.00	207.0

Instrument :
MSVOA_X
ClientSampleId :
BFB624

m/z	Abundance
36.10	750.0
37.00	4125.0
38.00	4094.0
39.10	1556.0
40.00	286.0
43.00	94.0
44.10	728.0
45.00	766.0
47.00	1344.0
47.90	592.0
49.10	3533.0
50.10	16552.0
51.00	4705.0
52.00	302.0
54.90	257.0
56.00	1261.0
57.00	2307.0
58.10	168.0
60.00	846.0
61.00	3232.0
62.10	3402.0
63.00	2408.0
64.00	338.0
66.80	197.0
68.00	8572.0
69.00	8019.0
70.20	693.0
71.90	526.0
73.00	3810.0
74.00	12857.0
75.00	41184.0
76.00	3769.0
77.00	379.0
77.80	268.0
78.10	255.0
78.90	2646.0
79.90	633.0
80.90	2440.0
81.90	535.0
86.10	72.0
86.90	3383.0
87.90	3267.0
91.00	246.0
92.00	2269.0
93.00	3110.0
94.00	9309.0
95.00	80392.0
96.10	5944.0
97.00	202.0
103.80	276.0
104.80	100.0
105.90	362.0
107.20	74.0
110.70	54.0
111.90	62.0
112.80	79.0
115.00	80.0
115.90	309.0
116.80	551.0
118.00	312.0
119.00	405.0
127.90	256.0
129.20	66.0
130.00	292.0
130.90	132.0
134.90	81.0
136.90	165.0
140.90	749.0
142.90	960.0
144.80	127.0
145.60	62.0
148.00	143.0
149.00	58.0
149.80	69.0
153.80	60.0
155.00	169.0
157.00	179.0
159.10	86.0
160.60	52.0
171.90	63.0
173.00	596.0
174.00	52184.0
175.00	3838.0
176.00	52000.0
177.00	3066.0
178.00	153.0

Instrument :
MSVOA_X
ClientSampleId :
BFB624

m/z	Abundance
36.00	990.0
37.10	5320.0
38.00	4430.0
39.10	1718.0
40.20	287.0
42.90	143.0
44.10	755.0
45.00	929.0
47.10	1079.0
48.00	717.0
49.00	4351.0
50.00	18528.0
51.00	6269.0
52.00	203.0
54.90	273.0
56.00	1693.0
57.00	3166.0
58.00	127.0
60.00	1067.0
61.00	5067.0
62.00	5243.0
63.00	3572.0
64.10	418.0
64.90	109.0
67.00	238.0
68.00	10402.0
69.00	10564.0
70.00	879.0
72.00	609.0
73.00	4369.0
74.00	18176.0
75.00	55512.0
76.00	5037.0
76.90	705.0
77.90	377.0
78.90	3378.0
79.90	790.0
80.90	2903.0
81.90	819.0
86.10	248.0
87.00	4423.0
88.00	4447.0
90.80	388.0
92.00	3223.0
93.00	4427.0
94.00	12963.0
95.00	108792.0
96.00	8207.0
97.00	278.0
104.00	460.0
105.00	130.0
105.90	425.0
107.00	85.0
109.80	69.0
110.70	96.0
111.80	54.0
114.70	148.0
116.00	389.0
116.80	604.0
117.80	387.0
119.00	659.0
127.90	346.0
129.10	218.0
129.90	335.0
130.90	62.0
134.90	285.0
136.90	237.0
139.10	67.0
140.90	1152.0
142.20	105.0
143.00	1163.0
145.80	110.0
147.90	284.0
148.90	52.0
149.90	103.0
153.90	73.0
155.10	239.0
156.90	206.0
159.00	176.0
160.70	142.0
172.10	173.0
173.00	936.0
174.00	80712.0
175.00	6125.0
176.00	75824.0
177.00	4752.0
178.00	178.0

Instrument :
MSVOA_X
ClientSampleId :
BFB624

m/z	Abundance
36.00	653.0
37.10	4072.0
38.10	3005.0
39.10	1046.0
40.00	269.0
44.10	540.0
45.00	725.0
47.10	882.0
47.90	480.0
49.00	3349.0
50.00	15027.0
51.10	4259.0
52.00	139.0
54.90	247.0
56.00	1379.0
57.00	2143.0
58.00	86.0
60.00	805.0
61.00	3746.0
62.00	3403.0
63.00	2403.0
64.00	294.0
64.80	60.0
67.10	270.0
68.00	7922.0
69.00	8330.0
69.90	804.0
72.00	476.0
73.00	3251.0
74.00	13534.0
75.00	41752.0
76.10	3599.0
77.00	344.0
78.00	469.0
79.00	2471.0
80.00	751.0
80.90	2706.0
81.90	451.0
85.80	73.0
86.20	105.0
87.00	3650.0
88.00	3246.0
90.90	276.0
92.00	2063.0
93.00	3701.0
94.00	10545.0
95.00	87184.0
96.00	5417.0
97.10	130.0
104.00	354.0
105.00	81.0
105.80	388.0
107.10	64.0
109.90	74.0
115.00	90.0
116.00	312.0
116.80	439.0
117.80	276.0
118.90	395.0
123.80	59.0
127.80	389.0
128.90	201.0
129.90	380.0
130.80	139.0
134.80	211.0
136.90	220.0
139.10	51.0
139.80	56.0
141.00	859.0
143.00	1075.0
145.80	97.0
147.90	214.0
149.00	59.0
149.70	160.0
152.70	62.0
153.90	53.0
154.80	249.0
157.00	118.0
158.80	166.0
160.90	107.0
171.80	162.0
173.00	705.0
174.00	74088.0
174.90	5571.0
176.00	72864.0
177.00	4687.0
178.10	112.0

Instrument :
MSVOA_X
ClientSampleId :
BFB624