

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041924\
 Data File : VX041171.D
 Acq On : 19 Apr 2024 12:32
 Operator : JC/MD
 Sample : P2198-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-05B-31-041824

Quant Time: Apr 20 01:29:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	189344	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	277078	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95960	46.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.880%
35) Dibromofluoromethane	5.385	113	84393	45.595	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.180%
50) Toluene-d8	8.647	98	315287	48.631	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.260%
62) 4-Bromofluorobenzene	11.079	95	107496	43.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.360%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1805	2.101	ug/l	99
24) 1,1-Dichloroethane	3.605	63	20856	6.882	ug/l	94
30) Chloroform	5.087	83	7841	2.381	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041924\
Data File : VX041171.D
Acq On : 19 Apr 2024 12:32
Operator : JC/MD
Sample : P2198-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-05B-31-041824

Quant Time: Apr 20 01:29:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

