

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041924\
 Data File : VX041177.D
 Acq On : 19 Apr 2024 14:50
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 20 01:31:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 02:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	255791	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	368075	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164819	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	125180	45.328	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.660%
35) Dibromofluoromethane	5.379	113	111491	45.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.680%
50) Toluene-d8	8.647	98	416048	48.307	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.079	95	146834	44.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041924\
Data File : VX041177.D
Acq On : 19 Apr 2024 14:50
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 20 01:31:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041624W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 02:04:44 2024
Response via : Initial Calibration

