

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042020\
 Data File : VX015759.D
 Acq On : 20 Apr 2020 12:36
 Operator : JC/SP
 Sample : VX0420WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0420WBL01

Quant Time: Apr 20 15:11:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1076801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1642204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1549434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	837340	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	624314	41.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.22%	
35) Dibromofluoromethane	5.47	113	500019	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	8.70	98	1929881	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.12	95	812802	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
Target Compounds						
17) Carbon Disulfide	2.55	76	143571	4.629	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042020\
Data File : VX015759.D
Acq On : 20 Apr 2020 12:36
Operator : JC/SP
Sample : VX0420WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0420WBL01

Quant Time: Apr 20 15:11:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

