

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042020\
 Data File : VX015770.D
 Acq On : 20 Apr 2020 16:57
 Operator : JC/SP
 Sample : PB128337TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128337TB

Quant Time: Apr 21 08:27:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1043092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1605740	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1550787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	892992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	628390	42.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.42%	
35) Dibromofluoromethane	5.47	113	501523	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	8.70	98	1925237	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.12	95	822321	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
Target Compounds						
16) Acetone	2.42	43	47317	7.486	ug/l	97
17) Carbon Disulfide	2.55	76	318253	10.594	ug/l	100
18) Methyl Acetate	2.75	43	17913	1.110	ug/l #	91
43) Isopropyl Acetate	6.42	43	398554	12.256	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	56165	1.124	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042020\
Data File : VX015770.D
Acq On : 20 Apr 2020 16:57
Operator : JC/SP
Sample : PB128337TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128337TB

Quant Time: Apr 21 08:27:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

