

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042020\
 Data File : VX015777.D
 Acq On : 20 Apr 2020 19:36
 Operator : JC/SP
 Sample : L2342-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-17-16-COMP

Quant Time: Apr 21 08:46:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1083098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1675383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1614600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	908283	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	643367	42.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.24%	
35) Dibromofluoromethane	5.47	113	515055	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	8.70	98	1971288	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	11.12	95	845135	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
Target Compounds						
16) Acetone	2.42	43	120758	18.400	ug/l	99
17) Carbon Disulfide	2.55	76	141261	4.528	ug/l	99
20) Methylene Chloride	2.84	84	65106	5.311	ug/l	93
43) Isopropyl Acetate	6.42	43	370994	10.934	ug/l	99
95) Naphthalene	13.82	128	73697	1.098	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042020\
Data File : VX015777.D
Acq On : 20 Apr 2020 19:36
Operator : JC/SP
Sample : L2342-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-17-16-COMP

Quant Time: Apr 21 08:46:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

