

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042020\
 Data File : VX015778.D
 Acq On : 20 Apr 2020 19:59
 Operator : JC/SP
 Sample : L2342-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-15-COMP

Quant Time: Apr 21 08:48:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1089175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1710791	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1645478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	916251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	653773	42.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.12%	
35) Dibromofluoromethane	5.47	113	528858	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	8.70	98	2026145	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.12	95	860258	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
Target Compounds						
16) Acetone	2.42	43	48055	7.281	ug/l	98
17) Carbon Disulfide	2.55	76	187252	5.969	ug/l	99
20) Methylene Chloride	2.83	84	38323	3.109	ug/l	89
43) Isopropyl Acetate	6.42	43	297452	8.585	ug/l	98
95) Naphthalene	13.82	128	697754	10.308	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042020\
Data File : VX015778.D
Acq On : 20 Apr 2020 19:59
Operator : JC/SP
Sample : L2342-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-15-COMP

Quant Time: Apr 21 08:48:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

