

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042020\
 Data File : VX015780.D
 Acq On : 20 Apr 2020 20:44
 Operator : JC/SP
 Sample : L2350-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5-D

Quant Time: Apr 21 08:52:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1091564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1663414	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1611907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	909322	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	644000	41.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.66%	
35) Dibromofluoromethane	5.47	113	514396	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	8.70	98	1988521	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.12	95	852748	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
Target Compounds						
16) Acetone	2.42	43	58932	8.910	ug/l	97
17) Carbon Disulfide	2.55	76	115181	3.664	ug/l	99
20) Methylene Chloride	2.84	84	97469	7.889	ug/l	99
43) Isopropyl Acetate	6.41	43	293556	8.714	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042020\
Data File : VX015780.D
Acq On : 20 Apr 2020 20:44
Operator : JC/SP
Sample : L2350-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5-D

Quant Time: Apr 21 08:52:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

