

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
 Data File : VX021763.D
 Acq On : 20 Apr 2021 19:05
 Operator : JC/MD
 Sample : M2064-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 21 11:23:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	154337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	258606	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	225389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	97003	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	87820	50.16	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.32%
35) Dibromofluoromethane	5.465	113	82078	49.63	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.26%
50) Toluene-d8	8.696	98	306672	48.23	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.46%
62) 4-Bromofluorobenzene	11.116	95	105466	44.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
Data File : VX021763.D
Acq On : 20 Apr 2021 19:05
Operator : JC/MD
Sample : M2064-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 21 11:23:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

