

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
 Data File : VX021766.D
 Acq On : 20 Apr 2021 20:15
 Operator : JC/MD
 Sample : PB135856TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135856TB

Quant Time: Apr 21 11:24:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	161430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	272174	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	233818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	103986	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	91408	49.91	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.82%		
35) Dibromofluoromethane	5.464	113	86199	49.52	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.04%		
50) Toluene-d8	8.696	98	317603	47.46	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.92%		
62) 4-Bromofluorobenzene	11.116	95	108546	43.43	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.86%		
Target Compounds						Qvalue
3) Chloromethane	1.319	50	3459	1.99	ug/l	99
16) Acetone	2.422	43	23492	37.31	ug/l	95
20) Methylene Chloride	2.837	84	10050	5.32	ug/l	93
25) 2-Butanone	4.635	43	10246	9.61	ug/l	98
43) Isopropyl Acetate	6.409	43	17788	4.50	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
Data File : VX021766.D
Acq On : 20 Apr 2021 20:15
Operator : JC/MD
Sample : PB135856TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135856TB

Quant Time: Apr 21 11:24:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

