

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
 Data File : VX021762.D
 Acq On : 20 Apr 2021 18:41
 Operator : JC/MD
 Sample : M2064-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 21 11:22:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	153407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	255741	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	223267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	94755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	88424	50.81	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.62%
35) Dibromofluoromethane	5.464	113	81882	50.07	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.14%
50) Toluene-d8	8.695	98	305668	48.61	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.22%
62) 4-Bromofluorobenzene	11.122	95	104673	44.57	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.14%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042021\
Data File : VX021762.D
Acq On : 20 Apr 2021 18:41
Operator : JC/MD
Sample : M2064-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 21 11:22:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.629 min Scan# 745
Delta R.T. -0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 153407
Ion Ratio Lower Upper
168 100
99 55.8 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 50.81 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41

Tgt Ion: 65 Resp: 88424
Ion Ratio Lower Upper
65 100
67 54.7 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.830 min Scan# 942
Delta R.T. -0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:	114	Resp:	255741
Ion	Ratio	Lower	Upper
114	100		
63	18.3	0.0	40.8
88	15.8	0.0	32.8



#35
Dibromofluoromethane
Concen: 50.07 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41

Tgt Ion:	113	Resp:	81882
Ion	Ratio	Lower	Upper
113	100		
111	101.6	81.9	122.9
192	20.5	16.7	25.1









#50
Toluene-d8
Concen: 48.61 ug/l
RT: 8.695 min Scan# 1248
Delta R.T. 0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 305668
Ion Ratio Lower Upper
98 100
100 65.5 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 44.57 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. 0.006 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41

Tgt Ion: 95 Resp: 104673
Ion Ratio Lower Upper
95 100
174 81.5 0.0 154.6
176 78.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. 0.001 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tgt Ion:117 Resp: 223267
Ion Ratio Lower Upper
117 100
82 54.3 47.4 71.2
119 32.3 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX021762.D
Acq: 20 Apr 2021 18:41

Tgt Ion:152 Resp: 94755
Ion Ratio Lower Upper
152 100
115 59.3 41.1 123.3
150 157.8 0.0 349.0

