

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028188.D
 Acq On : 20 Apr 2022 11:15
 Operator : JC/MD
 Sample : VX0420MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0420MBL01

Quant Time: Apr 20 15:17:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	403293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	661358	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	577833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	254780	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	228662	43.484	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.960%
35) Dibromofluoromethane	5.385	113	206758	44.579	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.160%
50) Toluene-d8	8.647	98	757013	42.496	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	85.000%
62) 4-Bromofluorobenzene	11.079	95	258045	37.885	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	75.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028188.D
Acq On : 20 Apr 2022 11:15
Operator : JC/MD
Sample : VX0420MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0420MBL01

Quant Time: Apr 20 15:17:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

