

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028192.D
 Acq On : 20 Apr 2022 13:02
 Operator : JC/MD
 Sample : N2468-09ME
 Misc : 4.05g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-15-1-8-GRME

Quant Time: Apr 21 06:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	448528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	749284	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	713986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	354569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	255436	43.677	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.360%		
35) Dibromofluoromethane	5.391	113	246756	46.959	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.920%		
50) Toluene-d8	8.653	98	874473	43.329	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	86.660%		
62) 4-Bromofluorobenzene	11.079	95	331411	42.947	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	85.900%		
Target Compounds					Qvalue	
16) Acetone	2.380	43	6433	2.218	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028192.D
Acq On : 20 Apr 2022 13:02
Operator : JC/MD
Sample : N2468-09ME
Misc : 4.05g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-15-1-8-GRME

Quant Time: Apr 21 06:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

