

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028197.D
 Acq On : 20 Apr 2022 15:02
 Operator : JC/MD
 Sample : N2448-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E14ST-B6-041522-GW-DUP

Quant Time: Apr 21 06:13:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	394960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	678355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	642495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	297546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289270	56.171	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.340%
35) Dibromofluoromethane	5.391	113	262609	55.202	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.400%
50) Toluene-d8	8.653	98	946721	51.814	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.620%
62) 4-Bromofluorobenzene	11.085	95	345988	49.524	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.040%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	8576	4.079	ug/l	96
30) Chloroform	5.099	83	72470	8.810	ug/l	100
47) Bromodichloromethane	7.824	83	8547	1.275	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028197.D
Acq On : 20 Apr 2022 15:02
Operator : JC/MD
Sample : N2448-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E14ST-B6-041522-GW-DUP

Quant Time: Apr 21 06:13:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

