

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028198.D
 Acq On : 20 Apr 2022 15:25
 Operator : JC/MD
 Sample : N2440-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20220411

Quant Time: Apr 21 06:13:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	387511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	666765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	634435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	296217	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	285820	56.568	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.140%
35) Dibromofluoromethane	5.391	113	255969	54.741	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.480%
50) Toluene-d8	8.653	98	941352	52.415	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.840%
62) 4-Bromofluorobenzene	11.079	95	339881	49.495	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028198.D
Acq On : 20 Apr 2022 15:25
Operator : JC/MD
Sample : N2440-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20220411

Quant Time: Apr 21 06:13:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

