

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028200.D
 Acq On : 20 Apr 2022 16:12
 Operator : JC/MD
 Sample : N2432-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB220414

Quant Time: Apr 21 06:14:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	375717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	648845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	606592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	280729	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275496	56.236	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 112.480%	
35) Dibromofluoromethane	5.391	113	251025	55.167	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 110.340%	
50) Toluene-d8	8.653	98	904291	51.742	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.480%	
62) 4-Bromofluorobenzene	11.079	95	326732	48.894	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028200.D
Acq On : 20 Apr 2022 16:12
Operator : JC/MD
Sample : N2432-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB220414

Quant Time: Apr 21 06:14:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

