

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028202.D
 Acq On : 20 Apr 2022 16:58
 Operator : JC/MD
 Sample : N2450-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 21 06:14:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	374283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	640123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	603589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	277384	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	267546	54.823	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.640%
35) Dibromofluoromethane	5.391	113	245903	54.777	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.560%
50) Toluene-d8	8.653	98	882662	51.193	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.380%
62) 4-Bromofluorobenzene	11.079	95	318371	48.292	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028202.D
Acq On : 20 Apr 2022 16:58
Operator : JC/MD
Sample : N2450-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 21 06:14:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

