

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028203.D
 Acq On : 20 Apr 2022 17:21
 Operator : JC/MD
 Sample : N2450-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 21 06:14:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	376092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	648957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	608700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	283038	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	272262	55.521	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.040%
35) Dibromofluoromethane	5.391	113	248087	54.512	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.020%
50) Toluene-d8	8.653	98	897802	51.362	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.720%
62) 4-Bromofluorobenzene	11.079	95	325706	48.732	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028203.D
Acq On : 20 Apr 2022 17:21
Operator : JC/MD
Sample : N2450-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 21 06:14:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

