

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028218.D
 Acq On : 21 Apr 2022 00:19
 Operator : JC/MD
 Sample : PB144231TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144231TB

Quant Time: Apr 21 06:20:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	381841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	663738	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	665620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	307201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	266818	53.591	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.180%			
35) Dibromofluoromethane	5.391	113	246120	52.875	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.760%			
50) Toluene-d8	8.653	98	914673	51.162	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.320%			
62) 4-Bromofluorobenzene	11.079	95	340589	49.824	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.640%			
Target Compounds						
16) Acetone	2.373	43	26739	16.279	ug/l	96
20) Methylene Chloride	2.788	84	200089	54.164	ug/l	99
43) Isopropyl Acetate	6.342	43	28429	2.835	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028218.D
Acq On : 21 Apr 2022 00:19
Operator : JC/MD
Sample : PB144231TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144231TB

Quant Time: Apr 21 06:20:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

