

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028222.D
 Acq On : 21 Apr 2022 01:52
 Operator : JC/MD
 Sample : PB144231ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144231ZHE#04

Quant Time: Apr 21 06:21:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	379632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	664392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	672237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	307182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	234108	47.295	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.580%		
35) Dibromofluoromethane	5.391	113	214995	46.143	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.280%		
50) Toluene-d8	8.653	98	799769	44.691	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	89.380%		
62) 4-Bromofluorobenzene	11.079	95	297981	43.548	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.100%		
Target Compounds						
16) Acetone	2.380	43	12006	6.582	ug/l	95
20) Methylene Chloride	2.794	84	19277	5.249	ug/l	94
43) Isopropyl Acetate	6.342	43	25932	2.584	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028222.D
Acq On : 21 Apr 2022 01:52
Operator : JC/MD
Sample : PB144231ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144231ZHE#04

Quant Time: Apr 21 06:21:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

