

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028226.D
 Acq On : 21 Apr 2022 03:25
 Operator : JC/MD
 Sample : PB144231ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144231ZHE#08

Quant Time: Apr 21 06:22:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	372937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	646789	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	654788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	291237	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	230824	47.469	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.940%
35) Dibromofluoromethane	5.385	113	212148	46.771	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.540%
50) Toluene-d8	8.653	98	783471	44.972	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.940%
62) 4-Bromofluorobenzene	11.079	95	287407	43.146	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.300%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	25788	16.057	ug/l	100
20) Methylene Chloride	2.788	84	192223	53.277	ug/l	97
43) Isopropyl Acetate	6.342	43	29104	2.979	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028226.D
Acq On : 21 Apr 2022 03:25
Operator : JC/MD
Sample : PB144231ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144231ZHE#08

Quant Time: Apr 21 06:22:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

