

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028228.D
 Acq On : 21 Apr 2022 04:11
 Operator : JC/MD
 Sample : PB144231ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144231ZHE#10

Quant Time: Apr 21 10:12:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	373886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	659833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	681027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	308942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	234288	48.059	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.120%		
35) Dibromofluoromethane	5.385	113	217775	47.062	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.120%		
50) Toluene-d8	8.653	98	807171	45.416	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	90.840%		
62) 4-Bromofluorobenzene	11.079	95	299891	44.130	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.260%		
Target Compounds						
16) Acetone	2.374	43	25056	15.518	ug/l	91
20) Methylene Chloride	2.788	84	196453	54.311	ug/l	98
43) Isopropyl Acetate	6.348	43	29073	2.917	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028228.D
Acq On : 21 Apr 2022 04:11
Operator : JC/MD
Sample : PB144231ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144231ZHE#10

Quant Time: Apr 21 10:12:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

