

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028187.D
 Acq On : 20 Apr 2022 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 15:16:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.323	0.283	12.4	109	0.00
3 P	Chloromethane	0.375	0.337	10.1	115	0.00
4 C	Vinyl Chloride	0.504	0.432	14.3#	106	0.00
5 T	Bromomethane	0.522	0.424	18.8	96	0.00
6 T	Chloroethane	0.461	0.318	31.0#	99	0.00
7 T	Trichlorofluoromethane	1.112	0.900	19.1	100	0.00
8 T	Diethyl Ether	0.382	0.312	18.3	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.443	-9.4	150#	0.00
10 T	Methyl Iodide	0.509	0.580	-13.9	130	0.00
11 T	Tert butyl alcohol	0.135	0.110	18.5	105	0.00
12 CM	1,1-Dichloroethene	0.384	0.409	-6.5#	145	0.00
13 T	Acrolein	0.080	0.069	13.7	88	0.00
14 T	Allyl chloride	0.645	0.588	8.8	110	0.00
15 T	Acrylonitrile	0.266	0.239	10.2	103	0.00
16 T	Acetone	0.215	0.200	7.0	121	0.00
17 T	Carbon Disulfide	1.035	1.012	2.2	121	0.00
18 T	Methyl Acetate	0.580	0.513	11.6	110	0.00
19 T	Methyl tert-butyl Ether	1.602	1.518	5.2	106	0.00
20 T	Methylene Chloride	0.484	0.480	0.8	126	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.483	7.6	107	0.00
22 T	Diisopropyl ether	1.469	1.288	12.3	101	0.00
23 T	Vinyl Acetate	1.283	1.142	11.0	100	0.00
24 P	1,1-Dichloroethane	0.867	0.795	8.3	112	0.00
25 T	2-Butanone	0.386	0.321	16.8	96	0.00
26 T	2,2-Dichloropropane	0.847	0.749	11.6	103	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.576	7.5	108	0.00
28 T	Bromochloromethane	0.372	0.319	14.2	96	0.00
29 T	Tetrahydrofuran	0.252	0.206	18.3	95	0.00
30 C	Chloroform	1.041	0.923	11.3#	102	0.00
31 T	Cyclohexane	0.794	0.684	13.9	99	0.00
32 T	1,1,1-Trichloroethane	0.943	0.853	9.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.545	16.4	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.351	0.321	8.5	101	0.00
36 T	1,1-Dichloropropene	0.457	0.417	8.8	102	0.00
37 T	Ethyl Acetate	0.464	0.403	13.1	96	0.00
38 T	Carbon Tetrachloride	0.516	0.496	3.9	104	0.00
39 T	Methylcyclohexane	0.577	0.537	6.9	103	0.00
40 TM	Benzene	1.339	1.241	7.3	101	0.00
41 T	Methacrylonitrile	0.239	0.209	12.6	97	-0.01
42 TM	1,2-Dichloroethane	0.497	0.446	10.3	99	0.00
43 T	Isopropyl Acetate	0.755	0.657	13.0	98	0.00
44 TM	Trichloroethene	0.388	0.384	1.0	110	0.00
45 C	1,2-Dichloropropane	0.328	0.295	10.1#	101	0.00
46 T	Dibromomethane	0.257	0.246	4.3	107	0.00
47 T	Bromodichloromethane	0.494	0.463	6.3	101	0.00
48 T	Methyl methacrylate	0.347	0.307	11.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028187.D
 Acq On : 20 Apr 2022 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 15:16:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	94	0.00
50 S	Toluene-d8	1.347	1.185	12.0	96	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.420	13.4	96	0.00
52 CM	Toluene	0.930	0.860	7.5#	100	0.00
53 T	t-1,3-Dichloropropene	0.525	0.506	3.6	101	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.528	6.4	101	0.00
55 T	1,1,2-Trichloroethane	0.378	0.358	5.3	105	0.00
56 T	Ethyl methacrylate	0.548	0.518	5.5	101	0.00
57 T	1,3-Dichloropropane	0.610	0.562	7.9	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.248	3.5	99	0.00
59 T	2-Hexanone	0.375	0.327	12.8	95	0.00
60 T	Dibromochloromethane	0.412	0.406	1.5	103	0.00
61 T	1,2-Dibromoethane	0.415	0.397	4.3	102	0.00
62 S	4-Bromofluorobenzene	0.515	0.452	12.2	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.391	0.397	-1.5	111	0.00
65 PM	Chlorobenzene	1.066	1.021	4.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.400	-0.8	106	0.00
67 C	Ethyl Benzene	1.788	1.752	2.0#	100	0.00
68 T	m/p-Xylenes	0.728	0.727	0.1	101	0.00
69 T	o-Xylene	0.725	0.709	2.2	100	0.00
70 T	Styrene	1.175	1.189	-1.2	101	0.00
71 P	Bromoform	0.313	0.325	-3.8	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.437	3.175	7.6	101	0.00
74 T	N-amyl acetate	1.188	1.065	10.4	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.004	10.0	100	0.00
76 T	1,2,3-Trichloropropane	0.984	0.917	6.8	100	0.00
77 T	Bromobenzene	0.895	0.832	7.0	103	0.00
78 T	n-propylbenzene	3.811	3.608	5.3	100	0.00
79 T	2-Chlorotoluene	2.369	2.133	10.0	98	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.796	4.1	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.305	7.9	96	0.00
82 T	4-Chlorotoluene	2.729	2.548	6.6	100	0.00
83 T	tert-Butylbenzene	2.979	2.835	4.8	102	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.785	4.0	100	0.00
85 T	sec-Butylbenzene	3.594	3.465	3.6	102	0.00
86 T	p-Isopropyltoluene	3.100	3.046	1.7	101	0.00
87 T	1,3-Dichlorobenzene	1.682	1.667	0.9	105	0.00
88 T	1,4-Dichlorobenzene	1.766	1.687	4.5	104	0.00
89 T	n-Butylbenzene	2.546	2.478	2.7	101	0.00
90 T	Hexachloroethane	0.498	0.511	-2.6	102	0.00
91 T	1,2-Dichlorobenzene	1.671	1.649	1.3	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.223	10.1	97	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.019	-2.5	111	0.00
94 T	Hexachlorobutadiene	0.443	0.425	4.1	108	0.00
95 T	Naphthalene	3.310	3.340	-0.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
Data File : VX028187.D
Acq On : 20 Apr 2022 10:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 20 15:16:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	0.995	-2.1	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6