



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042022\
 Data File : VX028213.D
 Acq On : 20 Apr 2022 21:37
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Title : SW846 8260
 Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	10483	PASS
75	95	30	60	48.1	31651	PASS
95	95	100	100	100.0	65755	PASS
96	95	5	9	6.8	4493	PASS
173	174	0.00	2	0.9	537	PASS
174	95	50	100	87.1	57272	PASS
175	174	5	9	7.7	4409	PASS
176	174	95	101	95.5	54704	PASS
177	176	5	9	6.5	3545	PASS