

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015790.D
 Acq On : 21 Apr 2020 13:37
 Operator : JC/SP
 Sample : PB128313TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128313TB

Quant Time: Apr 22 04:33:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1053471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1629927	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1572414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	856954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	626305	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	
35) Dibromofluoromethane	5.47	113	492906	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	8.70	98	1924285	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	11.12	95	799711	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
Target Compounds						
16) Acetone	2.42	43	50800	7.958	ug/l	99
17) Carbon Disulfide	2.56	76	66365	2.187	ug/l	97
43) Isopropyl Acetate	6.42	43	373329	11.310	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015790.D
Acq On : 21 Apr 2020 13:37
Operator : JC/SP
Sample : PB128313TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128313TB

Quant Time: Apr 22 04:33:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

